

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1390

Name: RITTER2  
Sterilizer ID: V2313228  
08/15/2022 10:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.7 F 0.0  
2:00 176.3 F 0.0  
4:00 188.5 F 1.1  
6:00 215.2 F 5.7  
8:00 223.1 F 10.4  
10:00 225.1 F 6.9  
12:00 237.4 F 11.5  
14:00 236.4 F 9.4  
16:00 238.7 F 9.7  
18:00 249.0 F 14.9  
20:00 260.5 F 21.4  
22:00 270.8 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.6 F 29.3  
2:00 271.6 F 28.5  
3:00 272.4 F 29.5  
4:00 272.3 F 29.0

Min 271.4 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.5 F 0.0  
5:00 235.8 F 0.0  
10:00 218.1 F 0.0  
15:00 212.4 F 0.0  
20:00 209.1 F 0.0  
25:00 209.3 F 0.0  
30:00 208.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 22:47  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:54  
CYCLE COMPLETE

08/15/2022 11:30 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_