
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000316
13:40 27/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:44
STERILIZING 5:04
136.0 C 228kPa 5:04
136.1 C 231kPa 5:34
136.2 C 231kPa 6:04
136.2 C 230kPa 6:34
136.4 C 233kPa 7:04
136.4 C 233kPa 7:34
136.3 C 231kPa 8:04
136.4 C 233kPa 8:34
136.4 C 233kPa 8:34
VENTING 8:34
AIR DRYING 9:26
CASSETTE REMOVED
CYCLE COMPLETE 14:39

Digital Signature #
2E7EFD76E31F1361
