

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2388

Name: RITTER 1  
Sterilizer ID: V2178198  
10/03/2022 12:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.3 F 0.0  
2:00 167.3 F 3.1  
4:00 182.4 F 6.8  
6:00 202.0 F 11.5  
8:00 218.8 F 5.6  
10:00 232.2 F 10.8  
12:00 230.5 F 7.1  
14:00 244.4 F 13.4  
16:00 243.1 F 11.9  
18:00 255.6 F 18.5  
20:00 267.0 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.1 F 28.7  
2:00 272.5 F 29.3  
3:00 271.6 F 28.5  
4:00 272.3 F 29.0  
  
Min 271.5 F 28.4  
Max 272.5 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.1 F 0.0  
5:00 233.9 F 0.0  
10:00 238.4 F 0.0  
15:00 232.9 F 0.0  
20:00 227.0 F 0.0  
25:00 230.2 F 0.0  
30:00 228.3 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 21:28  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:54  
CYCLE COMPLETE

10/03/2022 01:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_