
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001252
9:19 09/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:53
PRESSURIZING 3:56
STERILIZING 5:13
136.0 C 220kPa 5:13
136.3 C 224kPa 5:43
136.3 C 224kPa 6:13
136.2 C 222kPa 6:43
136.3 C 223kPa 7:13
136.1 C 221kPa 7:43
136.3 C 223kPa 8:13
136.4 C 225kPa 8:43
136.4 C 225kPa 8:43
VENTING 8:43
AIR DRYING 9:38
CYCLE COMPLETE 11:28

Digital Signature #
EA89DCCFB636F8BD
