
Midmark M11 - v1.0.5
Total Cycles: 3110

Name: RITTER2
Sterilizer ID: V2313228
11/10/2023 10:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.4 F 0.0
2:00 126.5 F 0.3
4:00 132.3 F 2.3
6:00 163.6 F 5.5
8:00 182.8 F 10.4
10:00 207.6 F 6.3
12:00 220.5 F 8.7
14:00 234.2 F 13.5
16:00 235.0 F 10.3
18:00 235.2 F 8.4
20:00 247.5 F 14.7
22:00 259.3 F 21.3
24:00 270.3 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.8 F 29.9
2:00 271.5 F 29.1
3:00 272.8 F 30.1
4:00 272.2 F 29.4

Min 271.3 F 29.1
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 211.8 F 0.0
10:00 195.3 F 0.0
15:00 193.9 F 0.0
20:00 190.6 F 0.0
25:00 189.3 F 0.0
30:00 188.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:52
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:02:60
CYCLE COMPLETE

11/10/2023 11:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____