

-----  
Midmark M11 - v1.0.5  
Total Cycles: 493

Name: RITTER2  
Sterilizer ID: V2313228  
10/27/2021 02:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.7 F 0.0  
2:00 179.5 F 3.8  
4:00 196.7 F 9.6  
6:00 219.4 F 6.2  
8:00 231.8 F 12.1  
10:00 237.5 F 10.7  
12:00 238.0 F 10.1  
14:00 252.6 F 17.4  
16:00 265.1 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.4  
1:00 272.1 F 29.8  
2:00 271.6 F 29.4  
3:00 271.5 F 29.5  
4:00 271.7 F 29.8  
  
Min 271.3 F 29.3  
Max 272.2 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.3 F 0.0  
5:00 238.5 F 0.0  
10:00 221.0 F 0.0  
15:00 214.8 F 0.0  
20:00 211.0 F 0.0  
25:00 197.5 F 0.0  
30:00 196.3 F 0.0  
Drying Complete

FILLING: 2:46  
HEATING: 17:32  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 00:57:15  
CYCLE COMPLETE

10/27/2021 03:08 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_