
Midmark M11 - v1.0.5
Total Cycles: 3644

Name: RITTER2
Sterilizer ID: V2313228
05/24/2024 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.3 F 0.0
2:00 108.7 F 0.2
4:00 121.7 F 1.8
6:00 151.9 F 4.7
8:00 182.1 F 8.1
10:00 202.2 F 11.8
12:00 212.8 F 5.6
14:00 221.6 F 10.0
16:00 229.5 F 8.6
18:00 237.0 F 11.4
20:00 235.7 F 9.4
22:00 248.7 F 15.5
24:00 259.7 F 21.7
26:00 270.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.5 F 30.1
2:00 271.7 F 29.4
3:00 271.5 F 29.9
4:00 272.3 F 29.9

Min 271.1 F 29.2
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 203.0 F 0.0
10:00 203.7 F 0.0
15:00 196.6 F 0.0
20:00 187.7 F 0.0
25:00 187.1 F 0.0
30:00 189.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:60
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:05:24
CYCLE COMPLETE

05/24/2024 02:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____