

-----  
Midmark M11 - v1.0.0  
Total Cycles: 637

Name: MIDMARK  
Sterilizer ID: V2178198  
01/27/2021 02:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.7 F 0.0  
2:00 157.4 F 2.5  
4:00 175.2 F 6.4  
6:00 197.0 F 11.7  
8:00 221.2 F 6.8  
10:00 235.7 F 12.5  
12:00 234.2 F 9.6  
14:00 234.4 F 8.3  
16:00 249.1 F 14.5  
18:00 261.4 F 21.4  
20:00 272.5 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.7 F 28.9  
2:00 271.3 F 28.8  
3:00 272.6 F 28.9  
4:00 271.7 F 28.6  
  
Min 271.3 F 28.6  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.5 F 0.0  
5:00 225.8 F 0.0  
10:00 226.8 F 0.0  
15:00 235.1 F 0.0  
20:00 226.7 F 0.0  
25:00 232.8 F 0.0  
30:00 226.4 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 20:52  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 00:59:12  
CYCLE COMPLETE

01/27/2021 03:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_