

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3903

Name: RITTER 1  
Sterilizer ID: V2178198  
10/13/2023 02:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.7 F 0.0  
2:00 157.4 F 3.2  
4:00 173.1 F 7.0  
6:00 196.1 F 11.5  
8:00 216.4 F 4.5  
10:00 226.9 F 9.3  
12:00 235.7 F 9.8  
14:00 240.4 F 11.4  
16:00 237.9 F 9.4  
18:00 251.3 F 15.9  
20:00 262.7 F 22.5  
22:00 272.8 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.5 F 29.1  
2:00 272.0 F 28.5  
3:00 271.5 F 28.6  
4:00 272.6 F 29.1

Min 271.2 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 205.1 F 0.0  
10:00 233.1 F 0.0  
15:00 227.8 F 0.0  
20:00 223.3 F 0.0  
25:00 219.4 F 0.0  
30:00 217.7 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 22:21  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:01:06  
CYCLE COMPLETE

10/13/2023 03:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_