
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000879
12:43 20/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 3:56
STERILIZING 5:15
136.0 C 228kPa 5:15
136.4 C 233kPa 5:45
136.4 C 233kPa 6:15
136.4 C 232kPa 6:45
136.4 C 233kPa 7:15
136.3 C 232kPa 7:45
136.3 C 232kPa 8:15
136.2 C 232kPa 8:45
136.2 C 232kPa 8:45
VENTING 8:45
AIR DRYING 9:44
CYCLE COMPLETE 24:44

Digital Signature #
0574EA29BAE70E50
