

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4386

Name: RITTER2  
Sterilizer ID: V2313228  
02/27/2025 11:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.0 F 0.0  
2:00 71.6 F 0.2  
4:00 75.5 F 0.8  
6:00 108.4 F 2.6  
8:00 169.9 F 5.5  
10:00 187.4 F 9.0  
12:00 201.7 F 12.6  
14:00 206.9 F 3.5  
16:00 218.9 F 6.7  
18:00 227.1 F 10.1  
20:00 236.6 F 12.5  
22:00 233.9 F 9.7  
24:00 240.1 F 10.7  
26:00 242.9 F 12.5  
28:00 253.7 F 18.1  
30:00 263.3 F 23.9  
32:00 272.5 F 30.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.4  
1:00 272.0 F 30.2  
2:00 271.9 F 29.8  
3:00 272.5 F 30.1  
4:00 272.5 F 30.4  
  
Min 270.9 F 29.2  
Max 273.2 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.3 F 0.0  
5:00 217.8 F 0.0  
10:00 236.8 F 0.0  
15:00 168.3 F 0.0  
20:00 156.7 F 0.0  
25:00 158.2 F 0.0  
30:00 162.6 F 0.0  
  
Drying Complete

FILLING: 0:52  
HEATING: 32:30  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:11:08  
CYCLE COMPLETE

02/27/2025 12:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_