

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2810

Name: RITTER2  
Sterilizer ID: V2313228  
08/29/2023 11:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
2:00 180.3 F 0.5  
4:00 199.4 F 4.6  
6:00 217.1 F 8.7  
8:00 234.4 F 13.5  
10:00 233.3 F 9.7  
12:00 238.6 F 9.4  
14:00 244.5 F 12.9  
16:00 245.1 F 13.1  
18:00 257.2 F 19.6  
20:00 268.1 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.4 F 29.3  
2:00 271.3 F 29.1  
3:00 272.1 F 29.2  
4:00 271.3 F 29.1

Min 271.1 F 28.7  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 209.2 F 0.0  
10:00 224.2 F 0.0  
15:00 223.6 F 0.0  
20:00 219.9 F 0.0  
25:00 219.2 F 0.0  
30:00 219.3 F 0.0  
Drying Complete

FILLING: 2:07  
HEATING: 21:17  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:00:55  
CYCLE COMPLETE

08/29/2023 12:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_