

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1714

Name: RITTER2  
Sterilizer ID: V2313228  
11/14/2022 09:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.3 F 0.0  
2:00 78.4 F 0.0  
4:00 96.2 F 0.0  
6:00 117.2 F 0.0  
8:00 149.6 F 0.2  
10:00 184.4 F 0.5  
12:00 203.8 F 2.7  
14:00 215.4 F 6.0  
16:00 225.9 F 9.5  
18:00 236.9 F 13.4  
20:00 235.6 F 9.7  
22:00 239.5 F 9.9  
24:00 244.8 F 13.0  
26:00 243.2 F 11.9  
28:00 255.4 F 18.3  
30:00 266.2 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.5 F 29.2  
2:00 271.3 F 28.5  
3:00 272.5 F 29.3  
4:00 271.6 F 28.5

Min 271.2 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 202.8 F 0.0  
10:00 222.4 F 0.0  
15:00 226.2 F 0.0  
20:00 231.4 F 0.0  
25:00 231.5 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:42  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:09:39  
CYCLE COMPLETE

11/14/2022 10:40 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_