
Midmark M11 - v1.0.0
Total Cycles: 3703

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2023 09:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.7 F 0.0
2:00 166.6 F 2.6
4:00 170.6 F 6.0
6:00 211.6 F 10.7
8:00 218.9 F 4.6
10:00 227.8 F 9.8
12:00 232.1 F 8.5
14:00 242.0 F 11.8
16:00 238.9 F 10.0
18:00 252.5 F 16.6
20:00 264.0 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.2 F 28.8
2:00 271.4 F 28.6
3:00 272.1 F 28.6
4:00 271.2 F 28.5

Min 270.9 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 232.2 F 0.0
10:00 224.6 F 0.0
15:00 230.2 F 0.0
20:00 225.8 F 0.0
25:00 226.4 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:23
CYCLE COMPLETE

08/30/2023 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____