
Midmark M11 - v1.0.5
Total Cycles: 2521

Name: RITTER2
Sterilizer ID: V2313228
06/15/2023 11:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.2 F 0.0
2:00 173.6 F 3.4
4:00 189.8 F 7.9
6:00 209.0 F 12.5
8:00 223.1 F 9.0
10:00 234.4 F 10.1
12:00 240.0 F 12.4
14:00 240.4 F 11.4
16:00 253.9 F 18.3
18:00 265.8 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.2 F 30.2
2:00 272.9 F 30.2
3:00 272.2 F 29.8
4:00 271.5 F 29.4

Min 271.2 F 29.3
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 224.7 F 0.0
10:00 214.1 F 0.0
15:00 213.2 F 0.0
20:00 217.2 F 0.0
25:00 220.9 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

06/15/2023 12:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____