
Midmark M11 - v1.0.0
Total Cycles: 5070

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2024 05:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.7 F 0.0
2:00 149.7 F 0.5
4:00 166.9 F 2.9
6:00 191.9 F 7.4
8:00 214.7 F 12.6
10:00 224.3 F 7.3
12:00 235.6 F 12.9
14:00 237.4 F 10.3
16:00 235.7 F 8.2
18:00 250.0 F 15.5
20:00 262.1 F 22.3
22:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.1 F 28.9
2:00 271.7 F 28.9
3:00 271.9 F 28.7
4:00 271.6 F 28.9

Min 271.3 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 200.3 F 0.0
10:00 198.6 F 0.0
15:00 197.1 F 0.0
20:00 203.7 F 0.0
25:00 213.2 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 22:19
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:00:48
CYCLE COMPLETE

11/07/2024 06:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____