
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000405
9:55 26/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:18
STERILIZING 4:33
136.0 C 220kPa 4:33
136.4 C 233kPa 5:03
136.2 C 232kPa 5:33
136.4 C 233kPa 6:03
136.2 C 231kPa 6:33
136.2 C 230kPa 7:03
136.3 C 231kPa 7:33
136.2 C 231kPa 8:03
136.2 C 231kPa 8:03
VENTING 8:03
AIR DRYING 8:56
CYCLE COMPLETE 9:02

Digital Signature #
00E10187320A0676
