
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001525
11:08 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:29
STERILIZING 3:40
136.4 C 227kPa 3:40
136.4 C 227kPa 4:10
136.4 C 226kPa 4:40
136.4 C 227kPa 5:10
136.2 C 224kPa 5:40
136.5 C 227kPa 6:10
136.4 C 227kPa 6:40
136.4 C 226kPa 7:10
136.4 C 226kPa 7:10
VENTING 7:10
AIR DRYING 8:04
CYCLE COMPLETE 8:19

Digital Signature #
DE8289FFB6881B8D
