
Midmark M11 - v1.0.5
Total Cycles: 4107

Name: RITTER2
Sterilizer ID: V2313228
11/11/2024 03:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.1 F 0.0
2:00 87.0 F 0.0
4:00 104.2 F 0.0
6:00 131.4 F 0.2
8:00 166.0 F 2.1
10:00 191.5 F 5.6
12:00 204.7 F 10.2
14:00 221.0 F 7.5
16:00 228.1 F 9.1
18:00 238.1 F 11.7
20:00 239.4 F 11.5
22:00 237.8 F 10.0
24:00 250.6 F 16.4
26:00 262.2 F 23.0
28:00 272.7 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.3 F 29.7
2:00 271.9 F 29.8
3:00 272.2 F 29.4
4:00 272.5 F 30.1

Min 271.4 F 29.2
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 201.7 F 0.0
10:00 204.3 F 0.0
15:00 198.5 F 0.0
20:00 197.0 F 0.0
25:00 199.3 F 0.0
30:00 199.3 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 28:23
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:06:41
CYCLE COMPLETE

11/11/2024 04:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____