

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4669

Name: RITTER 1  
Sterilizer ID: V2178198  
06/21/2024 01:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 110.4 F 0.0  
2:00 113.4 F 1.3  
4:00 133.9 F 3.8  
6:00 160.7 F 6.8  
8:00 186.8 F 9.9  
10:00 203.0 F 9.3  
12:00 212.0 F 5.2  
14:00 221.9 F 8.9  
16:00 232.1 F 12.9  
18:00 233.2 F 8.7  
20:00 244.2 F 12.9  
22:00 242.7 F 11.7  
24:00 254.0 F 17.4  
26:00 264.2 F 23.4  
28:00 272.1 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.5 F 29.1  
2:00 271.7 F 28.4  
3:00 271.5 F 28.7  
4:00 272.5 F 29.0

Min 271.2 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 216.5 F 0.0  
10:00 214.9 F 0.0  
15:00 212.9 F 0.0  
20:00 215.0 F 0.0  
25:00 227.5 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 28:12  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:07:03  
CYCLE COMPLETE

06/21/2024 02:25 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_