
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002328
9:13 31/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:42
STERILIZING 4:53
136.4 C 222kPa 4:53
136.3 C 224kPa 5:23
136.3 C 222kPa 5:53
136.4 C 224kPa 6:23
136.1 C 217kPa 6:53
136.4 C 224kPa 7:23
136.3 C 224kPa 7:53
136.3 C 218kPa 8:23
136.3 C 218kPa 8:23
VENTING 8:23
AIR DRYING 9:17
CYCLE COMPLETE 9:34

Digital Signature #
598D4F8DD85DC67E
