
Midmark M11 - v1.0.5
Total Cycles: 4296

Name: RITTER2
Sterilizer ID: V2313228
01/24/2025 06:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.3 F 0.0
2:00 152.1 F 2.9
4:00 179.0 F 7.5
6:00 205.1 F 13.1
8:00 222.9 F 7.2
10:00 238.1 F 12.8
12:00 237.3 F 10.2
14:00 235.5 F 8.5
16:00 250.4 F 15.8
18:00 262.8 F 23.0
20:00 272.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.4 F 29.0
2:00 272.8 F 29.5
3:00 272.1 F 29.3
4:00 272.3 F 28.9

Min 271.7 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 189.5 F 0.0
10:00 191.9 F 0.0
15:00 194.4 F 0.0
20:00 194.6 F 0.0
25:00 194.7 F 0.0
30:00 193.5 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

01/24/2025 07:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____