
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000080
13:37 14/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:11
STERILIZING 4:20
136.0 C 222kPa 4:20
136.4 C 225kPa 4:50
136.4 C 225kPa 5:20
136.4 C 226kPa 5:50
136.4 C 224kPa 6:20
CYCLE ABORTED 6:34
STOP BUTTON PRESSED
NOT STERILE
END TIME 6:34

Digital Signature #
5B74313CF1654943
