
Midmark M11 - v1.0.5
Total Cycles: 1822

Name: RITTER2
Sterilizer ID: V2313228
12/08/2022 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.8 F 0.0
2:00 81.2 F 0.0
4:00 123.0 F 0.0
6:00 123.0 F 0.0
8:00 162.0 F 0.0
10:00 191.6 F 1.4
12:00 198.9 F 5.1
14:00 216.1 F 8.7
16:00 231.1 F 12.7
18:00 230.3 F 7.8
20:00 242.4 F 12.8
22:00 239.8 F 10.7
24:00 237.3 F 9.1
26:00 251.0 F 15.8
28:00 262.3 F 22.4
30:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.4 F 29.1
2:00 271.6 F 28.7
3:00 272.4 F 29.1
4:00 271.4 F 28.5

Min 271.4 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 220.7 F 0.0
10:00 228.8 F 0.0
15:00 236.4 F 0.0
20:00 227.9 F 0.0
25:00 226.6 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:23
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:08:13
CYCLE COMPLETE

12/08/2022 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____