
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000590
11:29 19/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:41
STERILIZING 3:58
136.5 C 220kPa 3:58
136.4 C 232kPa 4:28
136.2 C 229kPa 4:58
136.1 C 229kPa 5:28
136.4 C 232kPa 5:58
136.4 C 231kPa 6:28
136.3 C 231kPa 6:58
136.4 C 231kPa 7:28
136.4 C 231kPa 7:28
VENTING 7:28
AIR DRYING 8:19
CYCLE COMPLETE 23:19

Digital Signature #
68DF8CF985EA6872
