

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3592

Name: RITTER 1  
Sterilizer ID: V2178198  
07/27/2023 02:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.1 F 0.0  
2:00 172.7 F 0.8  
4:00 191.4 F 4.2  
6:00 210.0 F 9.4  
8:00 223.4 F 6.7  
10:00 233.0 F 10.0  
12:00 237.5 F 8.1  
14:00 245.4 F 13.5  
16:00 244.3 F 12.3  
18:00 257.5 F 19.3  
20:00 268.8 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 271.8 F 28.2  
2:00 272.4 F 28.6  
3:00 272.7 F 28.9  
4:00 271.8 F 28.6  
  
Min 271.5 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.7 F 0.0  
5:00 222.9 F 0.0  
10:00 231.7 F 0.0  
15:00 227.3 F 0.0  
20:00 228.4 F 0.0  
25:00 227.9 F 0.0  
30:00 222.7 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 21:03  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:08  
CYCLE COMPLETE

07/27/2023 03:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_