

-----  
Midmark M11 - v1.0.5  
Total Cycles: 441

Name: RITTER2  
Sterilizer ID: V2313228  
10/12/2021 02:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.2 F 0.0  
2:00 146.1 F 2.4  
4:00 160.2 F 5.9  
6:00 181.3 F 10.6  
8:00 209.6 F 5.0  
10:00 225.2 F 10.1  
12:00 230.1 F 8.4  
14:00 240.6 F 12.7  
16:00 240.5 F 11.4  
18:00 253.1 F 17.8  
20:00 264.3 F 24.5  
22:00 271.6 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.4  
1:00 272.7 F 29.9  
2:00 271.5 F 29.3  
3:00 272.6 F 30.0  
4:00 271.8 F 29.3

Min 271.5 F 29.2  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.3 F 0.0  
5:00 212.2 F 0.0  
10:00 202.8 F 0.0  
15:00 207.3 F 0.0  
20:00 208.2 F 0.0  
25:00 209.8 F 0.0  
30:00 212.9 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 22:00  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:11  
CYCLE COMPLETE

10/13/2021 06:17 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01