
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002497
9:26 22/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:46
STERILIZING 4:37
136.0 C 220kPa 4:37
136.2 C 220kPa 5:07
136.4 C 223kPa 5:37
136.2 C 221kPa 6:07
136.3 C 223kPa 6:37
136.3 C 223kPa 7:07
136.3 C 222kPa 7:37
136.2 C 222kPa 8:07
136.2 C 222kPa 8:07
VENTING 8:07
AIR DRYING 9:03
CYCLE COMPLETE 9:06

Digital Signature #
FDB21E7C6CBF2403
