

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2897

Name: RITTER2  
Sterilizer ID: V2313228  
09/18/2023 10:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.8 F 0.0  
2:00 150.0 F 0.0  
4:00 170.0 F 2.0  
6:00 191.1 F 6.1  
8:00 207.4 F 10.3  
10:00 230.2 F 7.6  
12:00 231.7 F 10.1  
14:00 235.2 F 8.8  
16:00 243.0 F 12.6  
18:00 241.7 F 11.4  
20:00 253.6 F 17.6  
22:00 264.1 F 23.9  
24:00 271.4 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.8 F 29.7  
2:00 272.6 F 29.8  
3:00 272.0 F 29.9  
4:00 271.5 F 29.5

Min 270.7 F 28.7  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.9 F 0.0  
5:00 231.2 F 0.0  
10:00 214.9 F 0.0  
15:00 218.4 F 0.0  
20:00 193.2 F 0.0  
25:00 223.0 F 0.0  
30:00 211.7 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 24:01  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:02:33  
CYCLE COMPLETE

09/18/2023 12:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_