
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002638
10:24 31/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:45
STERILIZING 3:52
136.4 C 226kPa 3:52
136.3 C 225kPa 4:22
136.1 C 223kPa 4:52
136.2 C 224kPa 5:22
136.4 C 226kPa 5:52
136.4 C 225kPa 6:22
136.4 C 224kPa 6:52
136.4 C 225kPa 7:22
136.4 C 225kPa 7:22
VENTING 7:22
AIR DRYING 8:11
CYCLE COMPLETE 24:20

Digital Signature #
924FF30088BAFF30
