
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004111
13:21 26/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:59
STERILIZING 2:59
136.4 C 223kPa 2:59
136.4 C 226kPa 3:29
136.3 C 224kPa 3:59
136.0 C 218kPa 4:29
136.1 C 222kPa 4:59
136.2 C 223kPa 5:29
136.4 C 224kPa 5:59
135.9 C 221kPa 6:29
135.9 C 221kPa 6:29
VENTING 6:29
AIR DRYING 7:23
CYCLE COMPLETE 7:57

Digital Signature #
68DBF941E8224D83
