
Midmark M11 - v1.0.0
Total Cycles: 3018

Name: RITTER 1
Sterilizer ID: V2178198
03/20/2023 10:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.0 F 0.0
2:00 140.6 F 0.9
4:00 157.4 F 2.8
6:00 186.8 F 7.0
8:00 206.9 F 11.5
10:00 219.1 F 5.7
12:00 231.5 F 10.9
14:00 231.2 F 7.4
16:00 244.4 F 12.9
18:00 243.9 F 12.2
20:00 256.6 F 18.9
22:00 268.1 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.5
3:00 272.5 F 29.2
4:00 271.7 F 28.5

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 221.8 F 0.0
10:00 222.7 F 0.0
15:00 223.8 F 0.0
20:00 228.7 F 0.0
25:00 230.4 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:17
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:21
CYCLE COMPLETE

03/20/2023 11:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____