
Midmark M11 - v1.0.5
Total Cycles: 525

Name: RITTER2
Sterilizer ID: V2313228
11/09/2021 07:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.4 F 0.0
2:00 74.6 F 0.0
4:00 77.8 F 0.0
6:00 117.3 F 0.0
8:00 147.7 F 0.2
10:00 183.7 F 0.4
12:00 205.0 F 3.1
14:00 221.1 F 5.9
16:00 218.9 F 9.4
18:00 227.6 F 13.2
20:00 232.1 F 9.4
22:00 240.0 F 10.8
24:00 242.8 F 12.7
26:00 241.9 F 11.7
28:00 253.8 F 17.8
30:00 264.5 F 24.2
32:00 271.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 273.0 F 29.5
2:00 271.6 F 29.0
3:00 273.0 F 29.5
4:00 271.4 F 28.8

Min 271.4 F 28.8
Max 273.3 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 203.2 F 0.0
10:00 209.9 F 0.0
15:00 207.7 F 0.0
20:00 205.5 F 0.0
25:00 206.8 F 0.0
30:00 214.5 F 0.0

Drying Complete

FILLING: 2:14
HEATING: 32:03
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:11:27
CYCLE COMPLETE

11/09/2021 08:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____