
Midmark M11 - v1.0.0
Total Cycles: 3141

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2023 07:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.7 F 0.0
2:00 80.1 F 0.0
4:00 90.2 F 0.0
6:00 122.8 F 0.0
8:00 157.2 F 0.0
10:00 192.7 F 0.0
12:00 209.0 F 3.4
14:00 222.2 F 6.3
16:00 230.2 F 9.5
18:00 240.2 F 13.2
20:00 234.8 F 9.1
22:00 243.1 F 11.2
24:00 243.2 F 12.0
26:00 240.6 F 10.4
28:00 253.3 F 16.7
30:00 264.3 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.5 F 28.6
2:00 271.4 F 28.0
3:00 272.6 F 28.7
4:00 271.8 F 28.0

Min 271.4 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 220.0 F 0.0
10:00 206.0 F 0.0
15:00 205.0 F 0.0
20:00 212.2 F 0.0
25:00 218.2 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:60
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:10:04
CYCLE COMPLETE

04/21/2023 08:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____