
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003337
9:39 30/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:01
STERILIZING 3:09
136.4 C 220kPa 3:09
136.4 C 225kPa 3:39
136.2 C 220kPa 4:09
136.4 C 222kPa 4:39
136.4 C 224kPa 5:09
136.2 C 222kPa 5:39
136.2 C 221kPa 6:09
136.4 C 222kPa 6:39
136.4 C 222kPa 6:39
VENTING 6:39
AIR DRYING 7:31
CYCLE COMPLETE 18:10

Digital Signature #
00AB9F117E7831AC
