

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005315  
9:12 04/04/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:48  
PRESSURIZING 1:52  
STERILIZING 2:57  
136.4 C 222kPa 2:57  
136.4 C 224kPa 3:27  
136.4 C 223kPa 3:57  
136.3 C 224kPa 4:27  
136.4 C 224kPa 4:57  
136.3 C 223kPa 5:27  
136.1 C 221kPa 5:57  
136.4 C 223kPa 6:27  
136.4 C 223kPa 6:27  
VENTING 6:27  
AIR DRYING 7:22  
CYCLE COMPLETE 8:23

-----  
Digital Signature #  
3E55A3AA5A0ACF48  
-----