

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1814

Name: RITTER 1  
Sterilizer ID: V2178198  
04/12/2022 09:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.2 F 0.0  
2:00 165.5 F 1.5  
4:00 174.2 F 5.2  
6:00 196.8 F 9.6  
8:00 213.9 F 6.4  
10:00 227.6 F 8.0  
12:00 238.8 F 13.0  
14:00 237.6 F 9.7  
16:00 237.4 F 8.9  
18:00 247.9 F 14.1  
20:00 259.6 F 20.5  
22:00 270.3 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.8 F 28.9  
2:00 271.9 F 28.2  
3:00 272.1 F 28.7  
4:00 272.7 F 28.8

Min 271.4 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
5:00 229.7 F 0.0  
10:00 223.2 F 0.0  
15:00 214.3 F 0.0  
20:00 208.9 F 0.0  
25:00 204.2 F 0.0  
30:00 197.7 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 22:53  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:15  
CYCLE COMPLETE

04/12/2022 10:57 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02