
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002176
7:49 29/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:38
STERILIZING 3:40
136.4 C 222kPa 3:40
136.1 C 222kPa 4:10
136.0 C 220kPa 4:40
136.4 C 224kPa 5:10
136.2 C 222kPa 5:40
136.1 C 223kPa 6:10
136.2 C 222kPa 6:40
136.4 C 223kPa 7:10
136.4 C 223kPa 7:10
VENTING 7:10
AIR DRYING 8:01
CYCLE COMPLETE 38:01

Digital Signature #
385842F1CF3B37BA
