
Midmark M11 - v1.0.0
Total Cycles: 1219

Name: RITTER 1
Sterilizer ID: V2178198
09/29/2021 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.9 F 0.0
2:00 157.9 F 2.5
4:00 174.4 F 6.5
6:00 198.3 F 11.6
8:00 217.8 F 6.2
10:00 229.1 F 11.7
12:00 234.6 F 8.6
14:00 240.4 F 10.2
16:00 245.9 F 13.1
18:00 258.4 F 19.8
20:00 269.8 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.4 F 29.0
2:00 272.7 F 29.4
3:00 271.5 F 28.7
4:00 272.6 F 29.3

Min 271.4 F 28.7
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 233.7 F 0.0
10:00 226.6 F 0.0
15:00 220.7 F 0.0
20:00 215.9 F 0.0
25:00 215.0 F 0.0
30:00 210.9 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:42
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:59:57
CYCLE COMPLETE

09/29/2021 12:27 PM
ACCEPTED: X
REJECTED:
Operator: U01