
Midmark M11 - v1.0.5
Total Cycles: 1184

Name: RITTER2
Sterilizer ID: V2313228
06/09/2022 02:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.3 F 0.0
2:00 140.0 F 1.3
4:00 159.3 F 4.5
6:00 186.4 F 8.7
8:00 217.3 F 13.4
10:00 220.1 F 7.2
12:00 233.0 F 11.7
14:00 231.9 F 8.3
16:00 242.7 F 12.1
18:00 243.8 F 12.6
20:00 255.5 F 18.9
22:00 266.4 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.7 F 29.9
2:00 272.6 F 29.4
3:00 271.3 F 29.0
4:00 272.5 F 29.9

Min 271.3 F 28.9
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.7 F 0.0
5:00 213.1 F 0.0
10:00 210.7 F 0.0
15:00 207.3 F 0.0
20:00 206.2 F 0.0
25:00 203.2 F 0.0
30:00 201.7 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:42
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:11
CYCLE COMPLETE

06/09/2022 03:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____