
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000341
10:43 15/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:47
STERILIZING 4:57
136.0 C 228kPa 4:57
136.4 C 233kPa 5:27
136.4 C 232kPa 5:57
136.2 C 231kPa 6:27
136.2 C 229kPa 6:57
136.3 C 232kPa 7:27
136.0 C 229kPa 7:57
136.4 C 233kPa 8:27
136.4 C 233kPa 8:27
VENTING 8:27
AIR DRYING 9:19
CYCLE COMPLETE 24:19

Digital Signature #
8C124A7F4DFE1C62
