

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1281

Name: RITTER 1  
Sterilizer ID: V2178198  
10/20/2021 09:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
2:00 193.1 F 1.7  
4:00 193.6 F 5.7  
6:00 205.7 F 10.4  
8:00 220.0 F 5.8  
10:00 230.9 F 9.3  
12:00 240.5 F 11.0  
14:00 241.6 F 11.7  
16:00 239.5 F 10.1  
18:00 252.5 F 16.6  
20:00 263.6 F 23.0  
22:00 272.2 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.8 F 29.0  
2:00 271.8 F 28.3  
3:00 272.2 F 28.8  
4:00 272.5 F 28.8

Min 271.5 F 28.1  
Max 273.0 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 215.2 F 0.0  
5:00 228.9 F 0.0  
10:00 216.8 F 0.0  
15:00 216.1 F 0.0  
20:00 217.2 F 0.0  
25:00 224.7 F 0.0  
30:00 225.8 F 0.0  
Drying Complete

FILLING: 1:49  
HEATING: 22:12  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:01:23  
CYCLE COMPLETE

10/20/2021 10:44 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01