
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002043
10:20 16/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:23
STERILIZING 3:21
136.5 C 225kPa 3:21
136.3 C 226kPa 3:51
136.3 C 225kPa 4:21
136.4 C 226kPa 4:51
136.4 C 222kPa 5:21
136.2 C 225kPa 5:51
136.3 C 226kPa 6:21
136.2 C 225kPa 6:51
136.2 C 225kPa 6:51
VENTING 6:51
AIR DRYING 7:41
CYCLE COMPLETE 17:21

Digital Signature #
EFA26A5B039C2AB6
