
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 008334
14:32 20/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 2:59
STERILIZING 4:01
136.4 C 221kPa 4:01
136.4 C 225kPa 4:31
136.2 C 223kPa 5:01
136.4 C 225kPa 5:31
136.4 C 221kPa 6:01
136.4 C 220kPa 6:31
136.2 C 219kPa 7:01
136.4 C 225kPa 7:31
136.4 C 225kPa 7:31
VENTING 7:31
AIR DRYING 8:18
CYCLE COMPLETE 16:50

Digital Signature #
DB3751D80FD0A735
