
Midmark M11 - v1.0.5
Total Cycles: 4260

Name: RITTER2
Sterilizer ID: V2313228
01/13/2025 12:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.6 F 0.0
2:00 138.7 F 0.0
4:00 140.3 F 0.2
6:00 176.2 F 0.6
8:00 197.2 F 3.4
10:00 214.3 F 7.2
12:00 224.3 F 11.6
14:00 229.2 F 7.8
16:00 241.2 F 13.1
18:00 240.8 F 11.5
20:00 239.6 F 10.6
22:00 252.6 F 17.1
24:00 264.0 F 23.8
26:00 271.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.5 F 29.6
2:00 271.5 F 28.7
3:00 272.1 F 29.7
4:00 271.7 F 29.1

Min 271.2 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 223.9 F 0.0
10:00 230.9 F 0.0
15:00 200.1 F 0.0
20:00 195.1 F 0.0
25:00 210.1 F 0.0
30:00 230.9 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:07
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:04:44
CYCLE COMPLETE

01/13/2025 01:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____