

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 007729  
14:20 12/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:06  
PRESSURIZING 3:08  
STERILIZING 4:11  
136.4 C 222kPa 4:11  
136.1 C 219kPa 4:41  
136.4 C 225kPa 5:11  
136.3 C 223kPa 5:41  
136.4 C 223kPa 6:11  
136.4 C 223kPa 6:41  
136.4 C 225kPa 7:11  
136.4 C 224kPa 7:41  
136.4 C 224kPa 7:41  
VENTING 7:41  
AIR DRYING 8:33  
CYCLE COMPLETE 8:45

-----  
Digital Signature #  
F9C863CD222048ED  
-----