
Midmark M11 - v1.0.5
Total Cycles: 286

Name: RITTER2
Sterilizer ID: V2313228
08/19/2021 06:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.8 F 0.0
2:00 75.7 F 0.0
4:00 81.6 F 0.1
6:00 116.7 F 0.9
8:00 153.1 F 4.2
10:00 181.3 F 8.1
12:00 213.2 F 12.5
14:00 218.9 F 5.7
16:00 229.8 F 9.8
18:00 233.5 F 9.7
20:00 239.0 F 11.3
22:00 237.7 F 9.7
24:00 251.6 F 16.5
26:00 263.5 F 23.5
28:00 271.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.3 F 29.1
2:00 272.4 F 29.7
3:00 271.8 F 28.7
4:00 272.6 F 29.5

Min 271.6 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 173.1 F 0.0
10:00 168.6 F 0.0
15:00 166.0 F 0.0
20:00 165.5 F 0.0
25:00 168.6 F 0.0
30:00 165.2 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 28:02
STERILIZING: 4:00
VENTING: 2:49
DRYING: 30:00
TOTAL CYCLE: 01:06:24
CYCLE COMPLETE

08/19/2021 07:48 AM
ACCEPTED: X
REJECTED:
Operator: U02