
Midmark M11 - v1.0.5
Total Cycles: 2746

Name: RITTER2
Sterilizer ID: V2313228
08/09/2023 02:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.5 F 0.0
2:00 189.8 F 4.0
4:00 206.4 F 9.8
6:00 221.4 F 7.4
8:00 234.0 F 11.3
10:00 234.3 F 9.2
12:00 238.2 F 9.7
14:00 248.4 F 15.0
16:00 260.4 F 21.8
18:00 271.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.6
2:00 271.4 F 29.2
3:00 272.6 F 29.7
4:00 271.5 F 29.0

Min 271.4 F 29.0
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 228.2 F 0.0
10:00 223.9 F 0.0
15:00 222.9 F 0.0
20:00 220.0 F 0.0
25:00 229.0 F 0.0
30:00 228.1 F 0.0

Drying Complete

FILLING: 2:29
HEATING: 18:36
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

08/09/2023 03:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____