
Midmark M11 - v1.0.5
Total Cycles: 452

Name: RITTER2
Sterilizer ID: V2313228
10/14/2021 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.0 F 0.0
2:00 170.9 F 3.6
4:00 194.2 F 8.7
6:00 214.9 F 8.7
8:00 227.2 F 9.6
10:00 232.7 F 9.0
12:00 243.2 F 13.0
14:00 243.8 F 13.0
16:00 256.7 F 19.8
18:00 268.0 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.6 F 30.0
2:00 271.5 F 29.3
3:00 272.4 F 30.1
4:00 271.8 F 29.4

Min 271.4 F 29.2
Max 272.6 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 210.0 F 0.0
10:00 211.9 F 0.0
15:00 212.7 F 0.0
20:00 215.5 F 0.0
25:00 215.5 F 0.0
30:00 214.1 F 0.0
Drying Complete

FILLING: 2:41
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:59:03
CYCLE COMPLETE

10/14/2021 12:50 PM
ACCEPTED: X
REJECTED:
Operator: U02