
Midmark M11 - v1.0.5
Total Cycles: 186

Name: RITTER2
Sterilizer ID: V2313228
07/08/2021 03:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.6 F 0.0
2:00 167.7 F 3.3
4:00 192.0 F 7.8
6:00 212.7 F 13.3
8:00 226.1 F 8.1
10:00 239.1 F 12.7
12:00 241.5 F 11.9
14:00 241.1 F 11.3
16:00 254.6 F 18.2
18:00 266.3 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.5 F 29.3
2:00 272.5 F 29.6
3:00 272.2 F 29.1
4:00 272.5 F 29.8

Min 271.7 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 227.0 F 0.0
10:00 225.7 F 0.0
15:00 229.4 F 0.0
20:00 213.5 F 0.0
Dry Cycle Aborted

FILLING: 2:51
HEATING: 19:34
STERILIZING: 4:00
VENTING: 3:12
DRYING: 23:20
TOTAL CYCLE: 00:52:56

07/08/2021 04:17 PM
ACCEPTED: X
REJECTED:
Operator: U02